



Customer Site:

BSH Recycling
Vicarage Farm Road
Fengate,
Lincolnshire, PE1 5TP

Certificate of Accuracy

3882

Date: 19th October 2021

Platform Type :	15m 8300s	No. of load points :	8
Indicator Type :	DFWL		
Serial No. :	100 365660		
Maximum Cap. :	50,000 Kg	Minimum Capacity (20 x e) :	400 Kg
Increment (e) :	20 Kg	Class :	III

ECCENTRIC LOADING

Accuracy at load points to be tested:

Max. Capacity = Nearest

No./load points - 1 practical weight

50,000 = 7,000 KG

(7)

Pass ☒ Fail ☐

Accuracy across load points (Expanded Display ☐)

oe	oe	oe	oe
oe	oe	oe	oe

Weigh Office

Accuracy of Instrument - Linearity & Hysteresis (Expanded display ☐)

Known wt.	Increasing reading	Decreasing reading	Error
1 000 Kg	1 000 Kg	1 000 Kg	oe
7,000 Kg	7,000 Kg	7,000 Kg	oe
25,040 Kg	25,040 Kg	25,040 Kg	oe
34,040 Kg	34,040 Kg	34,040 Kg	oe
44 040 Kg	44,040 Kg		oe

Pass ☒

Fail ☐

ROLLING TESTS		LOAD 44040 kg	(Expanded display ☐)			
END	MIDDLE	END	Pass ☒	END	MIDDLE	END
oe	oe	oe	Fail ☐	oe	oe	oe

Deviations recorded as fractions of (e) unless expanded display used and recorded in kg.

	TEST	Verification
Permissible errors:	0 ≤ m ≤ 500e	+/-1e
	500 ≤ m ≤ 2000e	+/-2e
	2000 ≤ m ≤ 10000e	+/-3e
		+/-0.5e
		+/-1e
		+/-1.5e

The weighing instrument detailed in this certificate has been tested with calibrated weights traceable to the National Weights & Measures Laboratory as detailed in Certificate

Undertaken by

Y Boddock

No. TR11209 dated 23/3/21