

# CERTIFICATE OF CALIBRATION

ISSUED: Fri 24/Apr/2026

CERTIFICATE NUMBER: R08N002731

PAGE 1 OF 2 PAGES

## BLAKE & BOUGHTON Industrial Weighing Specialists

Units 8 & 10  
Roman Way  
Thetford  
Norfolk  
IP24 1XB

Tel: 01842 751555

Approved Signatory: Jamie Chadwick

Signature:



2026-04-24 08:24:09

### Customer

Bourne Skip Hire and Recycling Ltd,  
Cherry Holt Road,  
Bourne,  
Lincolnshire,  
PE10 9LA

Contact Adam Moore

### Calibration Site

Bourne Skip Hire and Recycling Ltd (Peterborough),  
Vicarage Farm Road,  
Fengate,  
Peterborough,  
PE1 5TP

Equipment	Capacity	Division	Test Equipment Used
<b>Make</b> Diniargeo	<b>1</b> 50 000kg	20kg	TR13194
<b>Model</b> DFWL	<b>2</b>		TP0057
<b>Serial No</b> 0100365660	<b>3</b>		MU30640
<b>Customer Ref</b>	<b>4</b>		
<b>Location</b> Yard			

### Comments

#### Notes

The weighing equipment described above has been calibrated using weights traceable to National Standards and in accordance with the following procedures (where relevant). The results were recorded.

#### ENGINEER CHECKS

The engineer has made the following checks prior to calibration and recorded any deviation that may affect the results. i. Equipment available for duration of calibration ii. Operation and parameters iii. Environmental factors iv. Condition of the equipment under test

#### CERTIFICATES AND TOLERANCES

Blake and Boughton will record measurements taken over the equipment's range and provide a Calibration Certificate showing performance to a specified tolerance. In instances where the accuracy specification of equipment being tested/calibrated is unknown, the general acceptance criteria will be an accuracy level of +/- 0.1% of scale capacity or one division, if the weighing equipment has less than 1000 divisions.

#### LINEARITY

A series of weights were added to the centre of the load receptor. The reading at each load was recorded. In the case of equipment with a capacity in excess of 500 kg or with restricted platform sizes it may be necessary to use 'make-up' weights. This does not affect the validity of the test.

#### ECCENTRICITY TEST

A load of approximately 1/3 of the machine capacity was placed in the centre of the load receptor and the readings were recorded. The load was then placed at each pan support in turn and again at the centre, the readings were recorded. Lesser loads may be used to meet customers' requirements. For moisture analysers and small circular top pan balances, a load of 1/3 or greater of the capacity of the machine was placed on three points of the top pan and the readings were recorded. Lesser loads may be used to meet customers' requirements.

#### REPEATABILITY

The repeatability load was applied to the centre of the load receptor and the reading recorded. The repeatability load was removed and the reading recorded.

#### ACCURACY

The certificate issued under this service is based on readings taken at a particular point of time and a particular location, it does not guarantee the accuracy of the equipment at any future time. The interpretation of the results declared is the responsibility of the customer having regard to the nature of the machine's use.

# CERTIFICATE OF CALIBRATION

ISSUED: Fri 24/Apr/2026

CERTIFICATE NUMBER: R08N002731

PAGE 2 OF 2 PAGES

**Make** Diniargeo  
**Model** DFWL  
**Serial No** 0100365660  
**Range Calibrated** 43 440kg x 20kg  
**Tolerance** ±0.1%  
**Type of Calibration** As Found

**Date of Calibration** Wed 22/Apr/2026  
**Next Calibration Due** April 2027  
**Calibrator** Jamie Chadwick  
**Approved Signatory** Jamie Chadwick  
**Customer Ref**  
**Location** Yard

As Found Eccentricity Test				Nominal Load: 7 960kg	
Ref	Indicated Reading (kg)	Ref	Indicated Reading (kg)		
1	7 960	6	7 940		
2	7 960	7	7 980		
3	7 940	8	7 960		
4	7 940	9	7 980		
5	7 940	10	7 960		

As Found Linearity Test	
Nominal Load (kg)	Indicated Reading (kg)
0	0
5 000	5 000
10 000	10 000
20 000	20 000
30 000	30 000
43 440	43 440

As Left Eccentricity Test				Nominal Load: 7 960kg	
Ref	Indicated Reading (kg)	Ref	Indicated Reading (kg)		
1	7 960	6	7 940		
2	7 960	7	7 980		
3	7 940	8	7 960		
4	7 940	9	7 980		
5	7 940	10	7 960		

As Left Linearity Test	
Nominal Load (kg)	Indicated Reading (kg)
0	0
5 000	5 000
10 000	10 000
20 000	20 000
30 000	30 000
43 440	43 440

END OF CERTIFICATE